

Canadian Biosafety Symposium
Date: June 1, 2009
Place: Delta Halifax, Halifax Nova Scotia

NSF49 Certification of Gaseous Chlorine Dioxide for BSC Decontamination

Paul Lorchheim,
ClorDiSys Solutions, Inc., Lebanon, NJ

Henry Luftman and Michael Regits,
DRS Laboratories, Allentown, PA



Canadian Food
Inspection Agency

Agence canadienne
d'inspection des aliments



Public Health
Agency of Canada

Agence de santé
publique du Canada

BSC Decontamination

- Formaldehyde gas has been the standard method for >30 yrs
- National Sanitation Foundation holds the standard for BSC maintenance
- Until now, only formaldehyde considered “pre-validated”
- Study to confirm similar status to chlorine dioxide (CD)

NSF/ANSI 49 – 2002

Annex G

Recommended microbiological decontamination procedure

“Prior to decontamination with an alternative [note added: other than depolymerized paraformaldehyde] method (such as VHP), cycle parameters and validation of those parameters must be developed for each model and size of BSC.”

Alternative CD Methods – Fixed Quantity (Method 1)


- A predetermined mass of CD created, dependent upon BSC volume
- Similar concept to current use of paraformaldehyde
- CD gas concentration not constant
- While concentration monitored in study, would not be measured in typical field use.
- Single exposure time for given CD/volume ratio

Alternative CD Methods – Fixed Concentration (Method 2)

- Concentration of CD maintained at set value
- Requires real-time concentration monitoring
- Requires ability to compensate for CD loss
- Concentration independent of cabinet size

NSF Biological Indicator Criteria (for each trial)

- 6 pairs of BIs
 - 3 pairs in downstream side of exhaust HEPA filter (center and opposite corners)
 - 1 pair within positive pressure plenum
 - 1 pair beneath cabinet workspace
 - 1 pair in center of upstream (dirty) side of down flow HEPA
- Residual CD on BI's to be neutralized prior to culturing



BI Validation Criteria

Site for single trial

- Success = If either of 2 samples is negative
- Failure = Both samples test positive
- Statistical analysis:
 - Assume spore strip has 2.3×10^6 spores (typical)
 - Assume of 2 strips, 1 tests positive and 1 negative →
 - 95% probability “true” efficacy at least 5.8 log kill
 - 50% probability “true” efficacy at least 6.3 log kill

BI Validation Criteria (cont.)

■ Single Trial

- Pass = all 6 sites pass
- Conditional Pass = 5 sites pass
- Fail = 2 or more sites fail

■ Cabinet/Method Validation

- At least 3 trials
- Pass = 3 trials pass
- Pass = 3 trials have conditional or full passes, with failed sites not coinciding
- Trials may be repeated if cause for failure identified



Validation Experiments

Method 1 (fixed mass of CD)

- Experimenters – Micro-Clean
- BSC Manufacturer – NuAire
- Location – NuAire (Plymouth, MN)
- Conditions:
 - 60 – 75% RH
 - CD amount - 0.1 g/ft^3
 - Exposure time – 80 min

Method 1 Preparation

BI PLACEMENT




SEALED CABINET

LEAK TESTING

Method 1 - Preparation

BI PLACEMENT



BI RETRIEVAL



BI Analysis

REMOVAL OF BI FROM POUCH



TRANSFER TO THIOSULFATE



TRANSFER TO
GROWTH MEDIA

Results – Method 1

A2 -6' Console


A2 – 4' Bench Top

Trial #	1	3	5		2	4	6	13
Dose	3.1	3.9	3.1		3.5	3.1	3.5	3.9
Position								
1	2	1	1		1	1	0	0
2	2	1	1		2	1	0	0
3	1	0	2		2	0	0	0
4	1	1	1		1	0	0	0
5	1	2	0		0	0	0	0
6	1	1	0		0	0	1	0
	F	CP	CP		F	P	P	P

FAIL

PASS*

A2 Failure due to operator error. Alcohol is not sporidical



0, 1, 2 = # positive
Bls out of 2

Dose – mg-hr/L

Results – Method 1

B1 -4' Console


B2 – 6' Console

Trial #	7	9	11		8	10	12
Dose	4.4	4.7	5.0		2.0	2.5	2.6
Position							
1	2	0	0		0	0	0
2	0	1	0		0	0	0
3	2	2	0		0	0	0
4	0	1	0		0	0	0
5	0	0	0		0	0	0
6	0	1	0		0	0	0
	F	CP	P		P	P	P

FAIL

PASS

B1 Failure due to operator error. Alcohol is not sporicidal



0, 1, 2 = # positive
Bls out of 2

Method 1 – Issues

- Likely BI handling problem through Run #7
 - Used forceps to transfer BIs from thiosulfate
 - Performed positive controls first
 - “Sterilized” forceps in alcohol between BIs
- Repeated one run on A2-4' to replace “bad” run – successful
- Possible penetration issue for exhaust HEPA on B1 cabinet

Results – Method 1

Repeat Study

A2 -6' Console

B1 – 6' Console

Trial #	15	17	19		14	16	18	20
Dose	3.0	3.7	3.7		2.6	3.7	3.3	3.3
Position								
1	0	0	0		0	0	0	0
2	0	0	0		0	0	0	0
3	0	0	0		0	0	0	0
4	0	0	0		0	0	0	0
5	0	0	0		0	0	0	0
6	0	0	0		0	0	0	0
	P	P	P		P	P	P	P

PASS

PASS



0, 1, 2 = # positive
Bls out of 2


Dose – mg-hr/L

Validation Experiments


Method 2 (fixed concentration of CD)

- Experimenters – ClorDiSys Solutions
- BSC Manufacturer – Baker
- Location – Micro-Clean (PA)
- Conditions:
 - 70% RH
 - Two concentrations (24+ trials)
 - 4.8 mg/L for 40 min
 - 2.8 gm/L for 55 min

Method 2 - Preparation



PLACING BIs



GAS MANIFOLD

Results – Method 2

4.8 mg/L for 40 min


A1 -6' Console

A2 – 4' Console

Trial #	1	6	7		2	5	15
Dose	3.2	3.2	3.2		3.2	3.2	3.2
Position							
1	0	0	0		0	0	0
2	1	0	0		0	0	0
3	0	0	0		0	0	0
4	0	0	0		0	0	0
5	0	0	0		0	0	0
6	0	0	0		0	0	0
	P	P	P		P	P	P

PASS

PASS



0, 1, 2 = # positive
Bls out of 2

Dose – mg-hr/L

Results – Method 2

4.8 mg/L for 40 min

B1 -6' Console


B2 – 4' Console

Trial #	3	4	10	11		12	13	14	16
Dose	3.2	3.2	3.2	3.2		3.2	3.2	3.2	3.2
Position									
1	0	0	0	0		2	0	1	0
2	1	0	1	0		1	0	0	0
3	0	0	0	0		2	0	0	0
4	0	0	0	0		1	0	0	1
5	0	0	0	0		0	0	0	0
6	0	ND	0	0		2	0	0	0
	P	P	P	P		F	P	P	P

PASS

PASS

B2 Failure due to NO recirculation. External circulation added for B2 cabinets



0, 1, 2 = # positive
Blis out of 2

Results – Method 2

2.8 mg/L for 55 min

A1

A2

B1

B2

Trial #	8	9	27		23	24	29		17	18	19		18	20	21
Dose	2.8	2.8	2.8		2.8	2.8	2.8		2.8	2.8	2.8		2.8	2.8	2.8
Position															
1	0	0	0		0	0	0		0	0	0		0	0	0
2	0	0	0		0	0	0		0	0	0		0	0	0
3	0	0	0		0	1	0		0	0	0		0	0	0
4	0	0	0		0	0	0		0	0	0		0	0	0
5	0	0	0		0	0	1		1	0	0		0	0	0
6	0	0	0		0	0	0		0	0	0		0	0	0
	P	P	P		P	P	P		P	P	P		P	P	P

Results

- Method 1 successfully validated on A2 console, A2 bench top, B1 and B2 cabinets
 - External circulation loop for A and B cabinets (from above exhaust HEPA to work space)
- Method 2 successfully validated on A1, A2, B1 and B2 cabinets
 - External circulation loop for B cabinets only

NSF Considerations

- NSF Task Force accepted results of validation study
- A revision to Annex G of NSF/ANSI 49 which includes chlorine dioxide gas as an alternative to formaldehyde and protocols for both methods has been accepted and available
 - Protocols use values of CD concentration and exposure time greater than used in validation

Annex G NSF / ANSI 49-2008

- Humidity 60-75%
- Method 1 Fixed Mass
 - 0.13 g / cu ft
- Method 2 Fixed Concentration
 - 3mg/L for 60 min
 - 5mg/L for 45 min
 - B Cabinets must have external circulation



Cloridox -
Manual
Single/Half
Method 1

Cloridox -
Manual
Double Pumper
Method 1



Chem-CD
Method 1

Generators



Mini Chlorine Dioxide System
(MCS)
Method 1




Minidox-M
Method 2



Cloridox-GMP
Method 2

Acknowledgements

- Micro-Clean, Inc.
 - Nu-Aire, Inc.



Test BI Locations

Validation Study for Biological Safety Cabinets (BSCs)

- BSCs used to simultaneously
 - protect biological material inside from external contamination
 - protect user and lab from biological material being manipulated
- HEPA filtration, blower(s), plenums
- Found in medical, veterinary, pharmaceutical, and biotech labs