

Honeywell

CERTIFICATE OF ANALYSIS

Page 1 of 1

Honeywell Specialty Chemicals Seelze GmbH D-30918 Seelze

Telefon: +49 5137 999-353

26.10.2023 / COAH_01080853

Analyzed for:
Honeywell PMT Advanced Materials E&C
Offering Management
115 Tabor Road, Morris Plains
NJ 07950, USA

QC Release Date: 26.Oct.2023
Production date: 08.Sep.2023
Expiry Date: 12.Aug.2028

Article/Product: 34741

Lot: N251A

HYDRANAL-Methanol dry

The quality certificate is valid for the time of delivery

HYDRANAL-Methanol dry is a medium for volumetric one-component Karl Fischer titration to be used with HYDRANAL-Composite.
HYDRANAL-Methanol dry can also be used to refill evaporated methanol in the coulometric KF oven titration.

Parameter	Specification	QC results
Suitability (for KF-Titration)*1	test	test passed
Assay of water*2	max. 0.01 %	0.002 %
Number of samples (n)	minimum 2	3

*1 volumetric KF titration according to Test Method 01 HYDRANAL version 13. Nov. 2017 (including ISO 760 version 1978 and ASTM E 203 version 2016).

The KF suitability test was performed on the first bottle from the filling line.

*2 coulometric KF titration according to Test Method 02 HYDRANAL version 13.Nov.2017

The assay of water is the highest value from a series of measurements on (n) samples.

- The expiry date is based on the current knowledge and holds only for proper storage conditions in the originally closed flasks/ packages.
- We herewith confirm that the delivery is effected according to the technical delivery conditions agreed.
- Particular properties of the products or the suitability for a particular area of application are not assured.
- We guarantee a proper quality within our General Conditions of Sales.

Thomas Wendt

Honeywell Specialty Chemicals Seelze GmbH
Thomas Wendt

Supervisor HYDRANAL Center of Excellence
Wunstorfer Str. 40, D-30926 Seelze



Quality Testing Lab Accredited According to DIN EN ISO/IEC 17025

